

# MTB IgG

**Enzyme Immunoassay for the semi  
quantitative determination of  
IgG antibodies to  
Mycobacterium Tuberculosis  
in human serum and plasma**

- for "in vitro" diagnostic use only -



**DIA.PRO**

**Diagnostic Bioprobes Srl  
Via G. Carducci n° 27  
20099 Sesto San Giovanni  
(Milano) - Italy**

Phone +39 02 27007161

Fax +39 02 44386771

e-mail: [info@diapro.it](mailto:info@diapro.it)

REF MTBG.CE  
96 Tests

## MTB IgG

### A. INTENDED USE

Enzyme ImmunoAssay (ELISA) for the qualitative and/or semi quantitative determination of IgG antibodies to Mycobacterium tuberculosis. The kit may be used for the follow up of patients undergoing tuberculosis.

### B. INTRODUCTION

Mycobacterium tuberculosis (MTB) is a fastidious, slowly-growing, strictly aerobic bacterium with a complex cell wall composed of peptide-glycans and many complex long-chain lipids.

Tuberculosis remains one of the most common and deadly diseases throughout the world. In the past 10 years there has been a resurgence of tuberculosis in old-world countries, also due to new infections (HIV) and immigration.

In the diagnosis of Tuberculosis and in the follow-up of infected patients, ELISA for antibodies may be useful to provide information on the immunological status of the patient, in addition to Nucleic Acid Tests (or NATs) able to determine the presence of the bacterium itself.

### C. PRINCIPLE OF THE TEST

Microplates are coated with a chimeric recombinant antigen bearing the most immunogenic epitopes of MTB.

The solid phase is first treated with the diluted sample and anti MTB IgG are captured, if present, by the antigens coated on the microplate.

After washing out all the other components of the sample, in the 2<sup>nd</sup> incubation bound anti MTB IgG antibodies are detected by the addition of polyclonal specific anti hlgG antibodies, labeled with peroxidase (HRP).

The enzyme captured on the solid phase, acting on the substrate/chromogen mixture, generates an optical signal that is proportional to the amount of anti MTB IgG antibodies present in the sample.

IgG in the sample may therefore be semi quantitated in arbU/ml by means of its S/Co value and a calibration curve.

### D. COMPONENTS

Code MTBG.CE contains reagents for 96 tests.

#### 1. Microplate **MICROPLATE**

n° 1 microplates. 12 strips of 8 breakable wells coated with a chimeric recombinant antigen bearing the most immunodominant epitopes of MTB. Plates are sealed into a bag with desiccant.

#### 2. Negative Control **CONTROL -**

1x4.0ml/vial. Ready to use control. It contains 1% goat serum proteins, 10 mM Na-citrate buffer pH 6.0 +/-0.1, 0.5% Tween 20, 0.09% Na-azide and 0.045% ProClin 300 as preservatives. The negative control is olive green color coded.

#### 3. Positive Control **CONTROL +**

1x4.0ml/vial. Ready to use control. It contains 1% goat serum proteins, human antibodies positive to MTB, 10 mM Na-citrate buffer pH 6.0 +/-0.1, 0.5% Tween 20, 0.09% Na-azide 0.045% ProClin 300 as preservatives. The Positive Control is dark green color coded.

#### 5. Wash buffer concentrate **WASHBUF 20X**

1x60ml/bottle. 20x concentrated solution containing 0.045% ProClin 300 as preservative. Once diluted, the wash solution contains 10 mM phosphate buffer pH 7.0+/-0.2 and 0.05% Tween 20 .

#### 6. Enzyme Conjugate **CONJ**

1x16ml/vial. Ready to use and red colour coded reagent. It contains Horseradish Peroxidase conjugated goat polyclonal antibodies to human IgG, 5% BSA, 10 mM Tris buffer pH 6.8+/-0.1, 0.045% ProClin 300 and 0.02% gentamicine sulphate as preservatives.

#### 7. Chromogen/Substrate **SUBS TMB**

1x16ml/vial. Ready-to-use component. It contains 50 mM citrate-phosphate buffer pH 3.5-3.8, 4% dimethylsulphoxide, 0.03% tetra-methyl-benzidine or TMB and 0.02% hydrogen peroxide or H<sub>2</sub>O<sub>2</sub>.

**Note: To be stored protected from light as sensitive to strong illumination.**

#### 8. Assay Diluent **DILAS**

1x8ml/vial. 10 mM tris buffered solution pH 8.0 +/-0.1 containing 0.045% ProClin 300 for the pre-treatment of samples and controls in the plate, blocking interference.

#### 9. Sulphuric Acid **H<sub>2</sub>SO<sub>4</sub> 0.3 M**

1x15ml/vial. It contains 0.3 M H<sub>2</sub>SO<sub>4</sub> solution. Attention: Irritant (H315, H319; P280, P302+P352, P332+P313, P305+ P351+P338, P337+P313, P362+P363).

#### 10. Sample Diluent: **DILSPE**

1x50ml/vial. It contains 1% goat serum proteins, 10 mM Na-citrate buffer pH 6.0 +/-0.1, 0.5% Tween 20, 0.09% Na-azide and 0.045% ProClin 300 as preservatives. To be used to dilute the sample.

**Note: The diluent changes color from olive green to dark bluish green in the presence of sample.**

11. Plate sealing foils n° 2

12. Package insert n° 1

### E. MATERIALS REQUIRED BUT NOT PROVIDED

1. Calibrated Micropipettes (200ul and 10ul) and disposable plastic tips.
2. EIA grade water (bidistilled or deionised, charcoal treated to remove oxidizing chemicals used as disinfectants).
3. Timer with 60 minute range or higher.
4. Absorbent paper tissues.
5. Calibrated ELISA microplate thermostatic incubator capable to provide a temperature of +37°C.
6. Calibrated ELISA microwell reader with 450nm (reading) and with 620-630nm (blinking) filters.
7. Calibrated ELISA microplate washer.
8. Vortex or similar mixing tools.

### F. WARNINGS AND PRECAUTIONS

1. The kit has to be used by skilled and properly trained technical personnel only, under the supervision of a medical doctor responsible of the laboratory.

2. When the kit is used for the screening of blood units and blood components, it has to be used in a laboratory certified and qualified by the national authority in that field (Ministry of Health or similar entity) to carry out this type of analysis.

3. All the personnel involved in performing the assay have to wear protective laboratory clothes, talc-free gloves and glasses. The use of any sharp (needles) or cutting (blades) devices should be avoided. All the personnel involved should be trained in biosafety procedures, as recommended by the Center for Disease Control, Atlanta, U.S. and reported in the National Institute of Health's publication: "Biosafety in Microbiological and Biomedical Laboratories", ed. 1984.

4. All the personnel involved in sample handling should be vaccinated for HBV and HAV, for which vaccines are available, safe and effective.

5. The laboratory environment should be controlled so as to avoid contaminants such as dust or air-borne microbial agents, when opening kit vials and microplates and when performing the test. Protect the Chromogen/Substrate from strong light and avoid vibration of the bench surface where the test is undertaken.
6. Upon receipt, store the kit at 2.8°C into a temperature controlled refrigerator or cold room.
7. Do not interchange components between different lots of the kits. It is recommended that components between two kits of the same lot should not be interchanged.
8. Check that the reagents are clear and do not contain visible heavy particles or aggregates. If not, advise the laboratory supervisor to initiate the necessary procedures for kit replacement.
9. Avoid cross-contamination between serum/plasma samples by using disposable tips and changing them after each sample. Do not reuse disposable tips.
10. Avoid cross-contamination between kit reagents by using disposable tips and changing them between the use of each one. Do not reuse disposable tips.
11. Do not use the kit after the expiration date stated on the external container and internal (vials) labels.
12. Treat all specimens as potentially infective. All human serum specimens should be handled at Biosafety Level 2, as recommended by the Center for Disease Control, Atlanta, U.S. in compliance with what reported in the Institutes of Health's publication: "Biosafety in Microbiological and Biomedical Laboratories", ed. 1984.
13. The use of disposable plastic-ware is recommended in the preparation of the liquid components or in transferring components into automated workstations, in order to avoid cross contamination.
14. Waste produced during the use of the kit has to be discarded in compliance with national directives and laws concerning laboratory waste of chemical and biological substances. In particular, liquid waste generated from the washing procedure, from residuals of controls and from samples has to be treated as potentially infective material and inactivated before waste. Suggested procedures of inactivation are treatment with a 10% final concentration of household bleach for 16-18 hrs or heat inactivation by autoclave at 121°C for 20 min..
15. Accidental spills from samples and operations have to be adsorbed with paper tissues soaked with household bleach and then with water. Tissues should then be discarded in proper containers designated for laboratory/hospital waste.
16. The Sulphuric Acid is an irritant. In case of spills, wash the surface with plenty of water
17. Other waste materials generated from the use of the kit (example: tips used for samples and controls, used microplates) should be handled as potentially infective and disposed according to national directives and laws concerning laboratory wastes.

#### G. SPECIMEN: PREPARATION AND RECOMMANDATIONS

1. Blood is drawn aseptically by venepuncture and plasma or serum is prepared using standard techniques of preparation of samples for clinical laboratory analysis. No influence has been observed in the preparation of the sample with citrate, EDTA and heparin.
2. Avoid any addition of preservatives to samples; especially sodium azide as this chemical would affect the enzymatic activity of the conjugate, generating false negative results.
3. Samples have to be clearly identified with codes or names in order to avoid misinterpretation of results. When the kit is used for the screening of blood units, bar code labeling and electronic reading is strongly recommended.
4. Haemolysed (red) and visibly hyperlipemic ("milky") samples have to be discarded as they could generate false results. Samples containing residues of fibrin or heavy particles or microbial filaments and bodies should be discarded as they could give rise to false results.

5. Sera and plasma can be stored at +2°...+8°C in primary collection tubes for up to five days after collection. Do not freeze primary tubes of collection. For longer storage periods, sera and plasma samples, carefully removed from the primary collection tube, can be stored frozen at -20°C for at least 12 months. Any frozen samples should not be frozen/thawed more than once as this may generate particles that could affect the test result.
6. If particles are present, centrifuge at 2.000 rpm for 20 min or filter using 0.2-0.8µ filters to clean up the sample for testing.

#### H. PREPARATION OF COMPONENTS AND WARNINGS

A study conducted on an opened kit has not pointed out any relevant loss of activity up to 6 re-use of the device and up to 3 months.

##### Microplates:

Allow the microplate to reach room temperature (about 1 hr) before opening the container. Check that the desiccant is not turned to dark green, indicating a defect of storage.

In this case call Dia.Pro's customer service.

Unused strips have to be placed back into the aluminum pouch, in presence of desiccant supplied, firmly zipped and stored at +2°..8°C. When opened the first time, residual strips are stable till the indicator of humidity inside the desiccant bag turns from yellow to green.

##### Controls:

Ready to use. Mix well on vortex before use.

##### Wash buffer concentrate:

The 20x concentrated solution has to be diluted with EIA grade water up to 1200 ml and mixed gently end-over-end before use. As some salt crystals may be present into the vial, take care to dissolve all the content when preparing the solution.

In the preparation avoid foaming as the presence of bubbles could give origin to a bad washing efficiency.

**Note:** *Once diluted, the wash solution is stable for 1 week at +2..8° C.*

##### Enzyme conjugate:

Ready to use. Mix well on vortex before use.

Be careful not to contaminate the liquid with oxidizing chemicals, air-driven dust or microbes.

If this component has to be transferred use only plastic, possibly sterile disposable containers.

##### Chromogen/Substrate:

Ready to use. Mix well on vortex before use.

Be careful not to contaminate the liquid with oxidizing chemicals, air-driven dust or microbes.

Do not expose to strong illumination, oxidizing agents and metallic surfaces.

If this component has to be transferred use only plastic, possible sterile disposable container.

##### Assay Diluent:

Ready to use. Mix well on vortex before use.

##### Sulphuric Acid:

Ready to use. Mix well on vortex before use.

Attention: Irritant (H315; H319; P280; P302+P352; P332+P313; P305+P351+P338; P337+P313; P362+P363).

Legenda:

##### Warning H statements:

**H315** – Causes skin irritation.

**H319** – Causes serious eye irritation.

**Precautionary P statements:**

**P280** – Wear protective gloves/protective clothing/eye protection/face protection.

**P302 + P352** – IF ON SKIN: Wash with plenty of soap and water.

**P332 + P313** – If skin irritation occurs: Get medical advice/attention.

**P305 + P351 + P338** – IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

**P337 + P313** – If eye irritation persists: Get medical advice/attention.

**P362 + P363** – Take off contaminated clothing and wash it before reuse.

**Sample Diluent:**

Ready to use. Mix well on vortex before use.

**I. INSTRUMENTS AND TOOLS USED IN COMBINATION WITH THE KIT**

1. Micropipettes have to be calibrated to deliver the correct volume required by the assay and must be submitted to regular decontamination (household alcohol, 10% solution of bleach, hospital grade disinfectants) of those parts that could accidentally come in contact with the sample. They should also be regularly maintained in order to show a precision of 1% and a trueness of +/-2%. Decontamination of spills or residues of kit components should also be carried out regularly.
2. The ELISA incubator has to be set at +37°C (tolerance of +/- 0.5°C) and regularly checked to ensure the correct temperature is maintained. Both dry incubators and water baths are suitable for the incubations, provided that the instrument is validated for the incubation of ELISA tests.
3. The **ELISA washer** is extremely important to the overall performances of the assay. The washer must be carefully validated in advance, checked for the delivery of the right dispensation volume and regularly submitted to maintenance according to the manufacturer's instructions for use. In particular the washer, at the end of the daily workload, has to be extensively cleaned out of salts with deionized water. Before use, the washer has to be extensively primed with the diluted Washing Solution. The instrument weekly has to be submitted to decontamination according to its manual (NaOH 0.1 M decontamination suggested).  
5 washing cycles (aspiration + dispensation of 350ul/well of washing solution + 20 sec soaking = 1 cycle) are sufficient to ensure the assay with the declared performances. If soaking is not possible add one more cycle of washing. An incorrect washing cycle or salt-blocked needles are the major cause of false positive reactions.
4. Incubation times have a tolerance of ±5%.
5. The ELISA microplate reader has to be equipped with a reading filter of 450nm and with a second filter (620-630nm, mandatory) for blanking purposes. Its standard performances should be (a) bandwidth ≤ 10 nm; (b) absorbance range from 0 to ≥ 2.0; (c) linearity to ≥ 2.0; repeatability ≥ 1%. Blanking is carried out on the well identified in the section "Assay Procedure". The optical system of the reader has to be calibrated regularly to ensure that the correct optical density is measured. It should be regularly maintained according to the manufacturer 's instructions.
6. When using an ELISA automated work station, all critical steps (dispensation, incubation, washing, reading, data handling) have to be carefully set, calibrated, controlled and regularly serviced in order to match the values reported in the sections "Validation of Test" and "Assay Performances". The assay protocol has to be installed in the operating system of the unit and validated as for the washer and the reader. In addition, the liquid handling part of the station

(dispensation and washing) has to be validated and correctly set. Particular attention must be paid to avoid carry over by the needles used for dispensing and for washing. This must be studied and controlled to minimize the possibility of contamination of adjacent wells. The use of ELISA automated work stations is recommended for blood screening when the number of samples to be tested exceed 20-30 units per run.

7. When using automatic devices, in case the vial holder of the instrument does not fit with the vials supplied in the kit, transfer the solution into appropriate containers and label them with the same label peeled out from the original vial. This operation is important in order to avoid mismatching contents of vials, when transferring them. When the test is over, return the secondary labeled containers to 2..8°C, firmly capped.
8. Dia.Pro's customer service offers support to the user in the setting and checking of instruments used in combination with the kit, in order to assure compliance with the requirements described. Support is also provided for the installation of new instruments to be used with the kit.

**L. PRE ASSAY CONTROLS AND OPERATIONS**

1. Check the expiration date of the kit printed on the external label of the kit box. Do not use if expired.
2. Check that the liquid components are not contaminated by naked-eye visible particles or aggregates. Check that the Chromogen/Substrate is colorless or pale blue by aspirating a small volume of it with a sterile transparent plastic pipette. Check that no breakage occurred in transportation and no spillage of liquid is present inside the box. Check that the aluminum pouch, containing the microplate, is not punctured or damaged.
3. Dilute all the content of the 20x concentrated Wash Solution as described above.
4. Dissolve the Calibrator as described above.
5. Allow all the other components to reach room temperature (about 1 hr) and then mix as described.
6. Set the ELISA incubator at +37°C and prepare the ELISA washer by priming with the diluted washing solution, according to the manufacturers instructions. Set the right number of washing cycles reported in the specific section.
7. Check that the ELISA reader has been turned on at least 20 minutes before reading.
8. If using an automated workstation, turn it on, check settings and be sure to use the right assay protocol.
9. Check that the micropipettes are set to the required volume.
10. Check that all the other equipment is available and ready to use.
11. In case of problems, do not proceed further with the test and advise the supervisor.

**M. ASSAY PROCEDURE**

The assay has to be carried out according to what reported below, taking care to maintain the same incubation time for all the samples in testing.

**Automated assay:**

In case the test is carried out automatically with an ELISA system, we suggest to make the instrument aspirate 200 ul Sample Diluent and then 10 ul sample.

All the mixture is then carefully dispensed directly into the appropriate sample well of the microplate. Before the next sample is aspirated, needles have to be duly washed to avoid any cross-contamination among samples.

Do not dilute controls/calibrator as they are ready to use.

Dispense 200 ul controls/calibrator in the appropriate control/calibration wells.

**Important Note:** Visually monitor that samples have been diluted and dispensed into appropriate wells. This is simply achieved by checking that the color of dispensed samples has

turned to dark bluish-green while the color of the negative control has remained olive green.

For the next operations follow the operative instructions reported below for the Manual Assay.

It is strongly recommended to check that the time lap between the dispensation of the first and the last sample will be calculated by the instrument and taken into consideration by delaying the first washing operation accordingly.

#### Manual assay:

- Place the required number of Microwells in the microwell holder. Leave the 1<sup>st</sup> well empty for the operation of blanking.
- Dispense 200 ul of Negative Control in triplicate and 200 ul Positive Control in single in proper wells. Do not dilute Controls as they are ready to use !
- Add 200 ul of Sample Diluent (DILSPE) to all the sample wells; then dispense 10 ul sample in each properly identified well. Mix gently the plate, avoiding overflowing and contaminating adjacent wells, in order to fully disperse the sample into its diluent.

**Important note:** Check that the color of the Sample Diluent, upon addition of the sample, changes from light green to dark bluish green, monitoring that the sample has been really added.

- Dispense 50 ul Assay Diluent (DILAS) into all the controls/calibrator and sample wells. Check that the color of samples has turned to dark blue.
- Incubate the microplate for **45 min at +37°C**.

**Important note:** Strips have to be sealed with the adhesive sealing foil, supplied, only when the test is carried out manually. Do not cover strips when using ELISA automatic instruments.

- Wash the microplate with an automatic washer by delivering and aspirating 350ul/well of diluted washing solution as reported previously (section I.3).
- Pipette 100µl Enzyme Conjugate into each well, except the 1<sup>st</sup> blanking well, and cover with the sealer. Check that this red coloured component has been dispensed in all the wells, except A1.

**Important note:** Be careful not to touch the plastic inner surface of the well with the tip filled with the Enzyme Conjugate. Contamination might occur.

- Incubate the microplate for **45 min at +37°C**.
- Wash microwells as in step 6.
- Pipette 100µl Chromogen/Substrate mixture into each well, the blank well included. Then incubate the microplate at **room temperature (18-24°C) for 15 minutes**.

**Important note:** Do not expose to strong direct illumination. High background might be generated.

- Pipette 100µl Sulphuric Acid into all the wells using the same pipetting sequence as in step 10 to stop the enzymatic reaction. Addition of acid will turn the positive control and positive samples from blue to yellow.
- Measure the color intensity of the solution in each well, as described in section I.5, at 450nm filter (reading) and at 620-630nm (background subtraction, mandatory), blanking the instrument on A1.

#### Important notes:

- Ensure that no finger prints are present on the bottom of the microwell before reading. Finger prints could generate false positive results on reading.
- Reading has to be carried out just after the addition of the Stop Solution and anyway not any longer than 20 minutes after its addition. Some self oxidation of the chromogen can occur leading to high background.

#### N. ASSAY SCHEME

Method	Operations
Controls	200 ul
Samples (S)	200ul [DILSPE]+10ul S
Assay Diluent (DILAS)	50 ul
<b>1<sup>st</sup> incubation</b>	<b>45 min</b>
Temperature	+37°C
Wash step	n° 5 cycles with 20" of soaking OR n° 6 cycles without soaking
Enzyme conjugate	100 ul
<b>2<sup>nd</sup> incubation</b>	<b>45 min</b>
Temperature	+37°C
Wash step	n° 5 cycles with 20" of soaking OR n° 6 cycles without soaking
TMB/H <sub>2</sub> O <sub>2</sub>	100 ul
<b>3<sup>rd</sup> incubation</b>	<b>15 min</b>
Temperature	r.t.
Sulphuric Acid	100 ul
Reading OD	450nm/620-630nm

An example of dispensation scheme is reported below:

#### Microplate

	1	2	3	4	5	6	7	8	9	10	11	12
A	BLK	S4										
B	NC	S5										
C	NC	S6										
D	NC	S7										
E	PC	S8										
F	S1	S9										
G	S2	S10										
H	S3	S11										

Legenda: BLK = Blank NC = Negative Control  
PC = Positive Control Sn = Sample

#### O. INTERNAL QUALITY CONTROL

A check is carried out on the controls and the calibrator any time the kit is used in order to verify whether their OD450nm/620-630nm values are as expected and reported in the table below.

Check	Requirements
Blank well	< 0.100 OD450nm value
Negative Control (NC)	< 0.050 mean OD450nm value after blanking
Positive Control	≥ 0.500 OD450nm value

If the results of the test match the requirements stated above, proceed to the next section.  
If they do not, do not proceed any further and operate as follows:

Problem	Check
<b>Blank well</b> > 0.100 OD450nm	1. that the Chromogen/Substrate solution has not got contaminated during the assay
<b>Negative Control (NC)</b> > 0.050 OD450nm after blanking	1. that the washing procedure and the washer settings are as validated in the pre qualification study; 2. that the proper washing solution has been used and the washer has been primed with it before use; 3. that no mistake has been done in the assay procedure (dispensation of positive control instead of negative control); 4. that no contamination of the negative control or of their wells has occurred due to positive samples, to spills or to the enzyme conjugate; 5. that micropipettes haven't got contaminated with positive samples or with the enzyme conjugate 6. that the washer needles are not blocked or partially obstructed.
<b>Positive Control</b> < 0.500 OD450nm	1. that the procedure has been correctly executed; 2. that no mistake has been done in the distribution of controls (dispensation of negative control instead of positive control. In this case, the negative control will have an OD450nm value > 0.150, too. 3. that the washing procedure and the washer settings are as validated in the pre qualification study; 4. that no external contamination of the positive control has occurred.

Should these problems happen, after checking, report any residual problem to the supervisor for further actions.

**Important note:**

*The analysis must be done proceeding as the reading step described in the section M, point 12.*

**P. RESULTS**

**P.1 Qualitative Assay**

Calculate the mean OD450nm/620-630nm value of the Negative Control (NC) and then apply the following formula:

$$\text{CUT-OFF} = \text{NC} + 0.350$$

**P.2 Semi Quantitative Assay**

Calculate the Sample / Cut-Off value (or S/Co) for the Controls and for the samples. Assign the value of 0 arbU/ml to the Negative Control and the value of 100 arbU/ml to the Positive Control. Then on a linear millimetre paper draw a line between the Negative Control and the Positive Control values.

S/Co values of samples are then converted into arbU/ml by means of the curve, providing a semi-quantification of the IgG content in the sample.

**Q. INTERPRETATION OF RESULTS**

In the **Qualitative Method**, samples showing an OD 450nm value lower than the Cut-Off value are considered negative for anti MTB IgG.

Samples showing an OD450nm/620-630nm value higher than the Cut-Off value are considered positive for anti MTB IgG.

In the **Semi-quantitative Method**, a quantification of the IgG content in arbU/ml is possible for those samples that show an

OD450nm/620-630nm higher than the Cut-Off (or S/Co > 1); this provide the possibility for the clinician to follow up the immunological status of the tuberculosis patient.

**Important notes:**

1. Interpretation of results should be done under the supervision of the laboratory supervisor to reduce the risk of judgment errors and misinterpretations.
2. Tests for antibodies alone do not provide the clinician with a reliable diagnosis of tuberculosis. Additional tests (for example NATs) should be carried out.
3. When test results are transmitted from the laboratory to another facility, attention must be paid to avoid erroneous data transfer.
4. Diagnosis has to be done and released to the patient by a suitably qualified medical doctor.

**R. PERFORMANCES**

Evaluation of Performances has been conducted on panels of positive and negative samples with reference to a CE marked reference kit.

**1. Diagnostic Sensitivity and Specificity:**

The Diagnostic **Sensitivity** was calculated on a panel of positive samples derived from patients with clinical evident sign of Tuberculosis, gently provided by a national centre for the control of Tuberculosis.

A value of > 95% was observed when referring to the reference device.

The Diagnostic **Specificity** was calculated on a panel of samples derived from normal people, without any history of Tuberculosis, and blood donors.

A value > 95% was observed.

These findings are summarized in the following table.

Sensitivity	> 95 %
Specificity	> 95 %

**2. Reproducibility:**

A study conducted on three samples of different anti M.tuberculosis IgG reactivity, examined in 16 replicates in three separate runs has shown CV% values ranging 10-20% depending on the OD450nm/620-630nm readings.

The variability shown in the tables did not result in sample misclassification.

**S. LIMITATIONS**

Frozen samples containing fibrin particles or aggregates may generate false positive results.

ELISA for the detection of antibodies to M.tuberculosis are recognized by the medical literature to provide only a limited information about the status of a patients and require other tests for a more reliable diagnosis of Tuberculosis.

**REFERENCES**

1. Ryan KJ; Ray CG (editors) (2004). Sherris Medical Microbiology, 4th ed., McGraw Hill. ISBN 0-8385-8529-9.
2. Cole ST; Brosch R; Parkhill J; et al. (1998). "Deciphering the biology of Mycobacterium tuberculosis from the complete genome sequence." Nature 393: 537-544.
3. Camus JC; Pryor MJ; Medigue C; Cole ST. (148). "Re-annotation of the genome sequence of Mycobacterium tuberculosis H37Rv". Microbiology 2002: 2967-2973.

4. Flowers T (1995). "Quarantining the noncompliant TB patient: catching the "Red Snapper"". Journal of health and hospital law : a publication of the American Academy of Hospital Attorneys of the American Hospital Association 28 (2): == h == 95-105. PMID 10141473.
5. Madigan, Michael; Martinko, John (editors) (2005). Brock Biology of Microorganisms, 11th ed., Prentice Hall. ISBN 0-13-144329-1.

All the IVD Products manufactured by the company are under the control of a certified Quality Management System according to ISO 13485 rule. Each lot is submitted to a quality control and released into the market only if conforming with the EC technical specifications and acceptance criteria.

Manufacturer:  
Dia.Pro Diagnostic Bioprobes Srl  
Via G. Carducci n° 27 – Sesto San Giovanni (MI) – Italy

